

## FFPE DNA Reference Material

**IMPORTANT:** Centrifuge tube prior to opening to prevent loss of product

### PRODUCT INFORMATION

Catalogue Number	HD-C1636
Batch ID	35891
Product Description	<a href="#">Idylla MSS FFPE Reference Standard</a>
Date of Manufacture	28-Jan-2019
Manufacturing QMS Status	ISO 13485

### PROPERTIES

Format	10 µm section
Storage Conditions	+4°C
Cell Density	100 x10 <sup>6</sup> cells per 1cm block
Use by	For the latest information on product shelf life please visit our website: <a href="https://www.horizondiscovery.com/reference-standards/coa-information">https://www.horizondiscovery.com/reference-standards/coa-information</a>

### CUSTOMER SUPPORT

Technical	<a href="mailto:technical@horizondiscovery.com">technical@horizondiscovery.com</a>
Orders	<a href="mailto:orders@horizondiscovery.com">orders@horizondiscovery.com</a>

### ADDITIONAL INFORMATION

Product is released under concession (NCM2019-0014)

#### Horizon Discovery

8100 Cambridge Research Park,  
Waterbeach, Cambridge,  
CB25 9TL, United Kingdom  
[www.horizondiscovery.com/](http://www.horizondiscovery.com/)

t + 44 (0) 1223 976 000

f + 44 (0) 1223 655 581

e

[orders@horizondiscovery.com](mailto:orders@horizondiscovery.com)

03-16-072 Rev. 00

**QUALITY CONTROL**

% of HCT116 cell line: Droplet Digital™ PCR

Expected % of HCT116 cell line	Measured % of HCT116 cell line	Acceptance Criteria	Pass/Fail
0%	0.032%	≤0.1%	PASS

Quality: Genomic DNA Screen Tape assay (Agilent Technologies)

Acceptance Criteria	DIN value	Pass/Fail
DIN value >7.0	8.9	PASS

**FOR RESEARCH USE ONLY**

Manufactured in the UK

**Horizon Discovery**

8100 Cambridge Research Park,  
Waterbeach, Cambridge,  
CB25 9TL, United Kingdom  
[www.horizondiscovery.com/](http://www.horizondiscovery.com/)

**t** + 44 (0) 1223 976 000

**f** + 44 (0) 1223 655 581

**e**

[orders@horizondiscovery.com](mailto:orders@horizondiscovery.com)