

FFPE DNA Reference Material

IMPORTANT: Centrifuge tube prior to opening to prevent loss of product

PRODUCT INFORMATION

Catalogue Number	HD-C1636
Batch ID	35892
Product Description	Idylla MSS FFPE Reference Standard
Date of Manufacture	28-Jan-2019
Manufacturing QMS Status	ISO 13485

PROPERTIES

Format	10 µm section
Storage Conditions	+4°C
Cell Density	100 x10 ⁶ cells per 1cm block
Use by	For the latest information on product shelf life please visit our website: https://www.horizondiscovery.com/reference-standards/coa-information

CUSTOMER SUPPORT

Technical	technical@horizondiscovery.com
Orders	orders@horizondiscovery.com

ADDITIONAL INFORMATION

Product is released under concession (NCM2019-0014)

Horizon Discovery

8100 Cambridge Research Park,
Waterbeach, Cambridge,
CB25 9TL, United Kingdom
www.horizondiscovery.com/

t + 44 (0) 1223 976 000

f + 44 (0) 1223 655 581

e

orders@horizondiscovery.com

03-16-072 Rev. 00

QUALITY CONTROL

% of HCT116 cell line: Droplet Digital™ PCR

Expected % of HCT116 cell line	Measured % of HCT116 cell line	Acceptance Criteria	Pass/Fail
0%	0.018%	≤0.1%	PASS

Quality: Genomic DNA Screen Tape assay (Agilent Technologies)

Acceptance Criteria	DIN value	Pass/Fail
DIN value >7.0	9.0	PASS

FOR RESEARCH USE ONLY

Manufactured in the UK

Horizon Discovery

8100 Cambridge Research Park,
Waterbeach, Cambridge,
CB25 9TL, United Kingdom
www.horizondiscovery.com/

t + 44 (0) 1223 976 000
f + 44 (0) 1223 655 581
e
orders@horizondiscovery.com